

## Calibration of FOTEMP Devices

**Weidmann Technologies Deutschland GmbH**

V1 2019-05-21

Tino Schurzmann

## Content

- 1 Calibration..... 3
  - 1.1 General..... 3
  - 1.2 Accuracy..... 4
- 2 Surface Measurements ..... 7
  - 2.1 Simulation..... 7
- 3 Sensor Replacement and general accuracy..... 8
  - 3.1 Measurement 1 - “Same Device - different Sensors” ..... 9

## 1 Calibration

### 1.1 General

Before a device leaves the rooms of Weidmann, a customer depending calibration is performed. This calibration results in a Calibration Table, unique for each device + sensor combination.

Depending on the calibration range several block calibrators are used.

We are equipped with:

- Dry-Block-Calibrator Ametek RTC-159-C
- Dry-Block-Calibrator Ametek ATC-250B
- Dry-Block-Calibrator Ametek RTC-700C
- PT100 measurement device Ametek DTI050
- heating source Omega 760
- calibration control software "Kalibrator2"




Figure 3 overview of calibration system

Figure 3 shows the calibration system. Liquid nitrogen is not shown there. This is one point where the sensor is let it and the device's response is included in the calibration table by hand.

The software "Kalibrator2" expecting a WEIDMANN thermometer and a calibrator connected. Now it is set up to step through a list of calibration temperatures (maximum is 16), given by the operator and before that from a customer.

The software executes at least two tasks: "calibration" and "adjustment" of a WEIDMANN thermometer. During the "calibration" step the spectrometer's response at the 16 calibration points is found and stored in the device. After running through all points, the software restarts at the former first step and measures the 16 temperatures again to see the device's differences. It stores them and the calibration table in text files for backup reasons and the operator to check and calculate the standard deviation of the device and to create the device's certificate.

This certificate is created by reading the software's text files (unchanged), checking the accuracy and generating the certificate (PDF) automatically.

An example of the calibration table is shown in Figure 4. It can be read and written even after the device left Weidmann. For not dealing with floating numbers the temperature is multiplied by ten. The EDGE value is one special value out of the gallium arsenide reflection spectrum, showing the temperature depending moved reflection spectrum. This value corresponds to the spectrometers output and the optical setup. Due to the manufacturing inaccuracies there is no way to neglect calibration at desired accuracy.

Temperature x 10	EDGE
1500	15801
1400	16115
1300	16417
...	...
-300	20307
-400	20509
-500	20701

Figure 4 Example of FOTEMP Calibration Table

## 1.2 Preinstallation of the PC

For up. And downloading calibration data the user can read the Device' interface documentetion

[http://www.optocon.de/fileadmin/media/fotemp/FotempCommunicationCompact\\_OEM\\_Rev\\_14.pdf](http://www.optocon.de/fileadmin/media/fotemp/FotempCommunicationCompact_OEM_Rev_14.pdf) from [www.optocon.de](http://www.optocon.de) or follow the steps in the next chapters of this document.

The later used scripts are Python scripts. Therefore the programming language interpreter "Python" has to be installed in its version 2.7.10 or higher. Additional a communication via a COM port (RS232 or USB) is needed, so the package "pyserial-2.7" must be installed too. This document shows the usage on a Win10 PC. Linux works in the same way.

All scripts need a COM port, we are starting here.

1. Click with your mouse on the lower left desktop corner and choose "Device manager" and find the COM port under "Connections"
2. download "python-2.7.16" from here for your system:  
<https://www.python.org/downloads/release/python-2716/>, Linux has Python preinstalled
3. download "pyserial-2.7.win32" from here for your system:  
<https://sourceforge.net/projects/pyserial/files/pyserial/2.7/>, Linux: enter "pip install pyserial" in bash
4. install Python (from step 1)
5. install "PySerial" from step 2
6. reboot your Windows PC
7. copy the script files (find them on [www.optocon.de](http://www.optocon.de))
8. only for Linux: change the lines "*ser = serial.Serial("COM"+str(comport), COMPORT\_SPEED, timeout=COMPORT\_TIMEOUT)*" to "*ser = serial.Serial("<your\_registered\_device\_path>"+str(comport), COMPORT\_SPEED, timeout=COMPORT\_TIMEOUT)*" in all scripts
9. open a command line (Win+R, enter "cmd", press Enter key, Ctrl+Alt+T on Linux)
10. browse to the script folder


You are now ready to execute the scripts.

## 1.3 Download and Upload of Calibration Data

The target of this section is to describe the process of downloading a calibration file from the device to a PC. Given you have a running Python inclusive PySerial and you have the scripts

A 4-channel OEM-PLUS device was set up and is running. The device was connected to a Win10 PC. "Device Manager" shows a connected device "USB Serial Port (COM8)", so our COM port is "8".

Starting a command line on Win10 means pressing WindowsKey+R together, entering "cmd" and pressing enter key. Changing the partition and change the directory to the downloaded scripts is shown in Figure 1.



```
C:\Windows\system32\cmd.exe
Microsoft Windows [Version 10.0.17763.475]
(c) 2018 Microsoft Corporation. Alle Rechte vorbehalten.

C:\Users\Tino>D:

D:\>cd SW_OC\calibration_package

D:\SW_OC\calibration_package>dir
Datenträger in Laufwerk D: ist DATA
Volumeseriennummer: 5810-0275

Verzeichnis von D:\SW_OC\calibration_package

21.05.2019  13:55    <DIR>          .
21.05.2019  13:55    <DIR>          ..
21.05.2019  13:54             1.571 calitab_read.py
21.05.2019  13:57             2.422 calitab_write.py
                2 Datei(en),          3.993 Bytes
                2 Verzeichnis(se), 805.712.736.256 Bytes frei

D:\SW_OC\calibration_package>
```

Figure 1 listing the script files

The reading script downloads or reads calibration data from a FOTEMP device (not ModuleSystem). It needs to know the COM port number and the channel to be read. After reading one channel it creates a text file with the device's calibration data in one file per channel for not mixing anything. The device's data are written in the text files "data\_read\_ch<1/2/3/4>.cal". Compare Figure 2 for reading channel "3".

The writing script writes calibration data from a text file to the FOTEMP device (not ModuleSystem). It needs to know the COM port number and the channel to write to. Always read the device's data before overwriting anything. When the COM port number and the channel is known, the script expects the data to be written in a file with a special file name as a safety action. For channel 1 the data must be in a text file called "data\_write\_ch1.cal", for channel 3 the data must be in "data\_write\_ch3.cal".

Use your favorite text editor to change these files for your needs.

```
C:\Windows\system32\cmd.exe

D:\SW_OC\calibration_package>python calitab_read.py
read calibration from FOTEMP

Which COM (number only)? 8
Which channel (1-8)? 3
used com port: COM8
---
14156 1500
14522 1370
14862 1250
15219 1120
15574 990
15894 870
16235 740
16578 610
16887 490
17217 360
17543 230
17836 110
18152 -20
18460 -150
18735 -270
19032 -400
data written to file <data_read_ch3.cal>

D:\SW_OC\calibration_package>
```

Figure 2 reading channel “3” of the device at COM8

```
C:\Windows\System32\cmd.exe

D:\SW_OC\calibration_package>python calitab_write.py
write calibration to FOTEMP

for writing to channel 1, the file <data_write_ch1.cal> is expected
for writing to channel 2, the file <data_write_ch2.cal> is expected
for writing to channel 3, the file <data_write_ch3.cal> is expected
for writing to channel 4, the file <data_write_ch4.cal> is expected
Which COM (number only)? 8
Which channel (1-8)? 4
used com port: COM8
executing :74 4 3383 5DC 34F7 55A 3647 4E2 37B2 460 3914 3DE 3A56 366 38B1 2E4 3D02 262 3E37 1EA 3F83 168 40C6 E6 41EC 6
E 4325 FFEC 4457 FF6A 456D FEF2 4695 FE70
done.

D:\SW_OC\calibration_package>
```

Figure 3 writing “data\_write\_ch4.cal” to the device at COM8